Rapid, sensitive and accurate method for determination of Lafutidine hydrochloride in human plasma by RP-HPLC system

Paras P. Vekariya, Hitendra S. Joshi*

Department of Chemistry, Saurashtra University, Rajkot - 36005, India

*E-mail address: drhsjoshi49@gmail.com

ABSTRACT

Simple and rapid reverse phase high-performance liquid chromatography (RP-HPLC) method was developed and validated using Phenomenex Gemini c18 (4.6 x 250 mm, 5 μ) reverse phase column for the determination of LAF in human plasma, Solid Phase Extraction (SPE) technique was used for the extraction of analyte, detection was carried out by Photo Diode Array detector at 216 nm. Chromatographic resolution of the LAF was achieved within 4.6 min by using mobile phase Methanol and 5 mM Di-Potassium Hydrogen Phosphate Buffer (pH 9.5) (80:20, v/v), flow rate was 1.0 mL/min. Calibration curve was linear with correlation coefficient of 0.9996 in the range of 50-1000 ng/mL, Limit of Detection (LOD) and Limit of Quantitation (LOQ) were 10 ng/mL and 30 ng/mL respectively, intra and inter-day deviations were lower than 3.92 % and 3.98 % respectively. The overall mean recovery of LAF was 94.57 %. No any endogenous constituents were found to interfere at retention time of the analyte. This new RP-HPLC method was successfully validated and may be applied to conduct bioavailability & bioequivalence studies of LAF.

Keywords: HPLC-PDA; Human plasma; Solid phase Extraction

1. INTRODUCTION

Lafutidine Hydrochloride (LAF) is 2-(furan-2-ylmethylsulfinyl)-N-[4-[4-(piperidin-1-ylmethyl) pyridin-2-yl] oxybut-2-enyl] acetamide refer (Figure 1) (Merck index, 2001) [1]. It belongs to the class of histamine 2 (H₂) -receptor antagonists and it used in the treatment of peptic ulcer and gastro-oesophageal reflux disease.

It is a second-generation H₂-receptor antagonist with an increased action on the gastric mucosal-defensive capacity. It has been shown to exhibit potent gastro protective activity in addition to gastric acid antisecretory effect [2-4].

Figure 1. Chemical structure of Lafutidine hydrochloride salt.

It is reported that to cure ulcer H₂-receptor therapy was used but after this therapy it was observed that reformation of ulcer was occur rapidly and frequently to prevent such kind of consequence gastroprotective H₂-receptor antagonist drug like famotidine were given along with h₂-receptor therapy and LAF is one H₂-receptor antagonist [5-7].

There are some methods which reports determination of LAF in human plasma and in serum by high-performance liquid chromatography electro spray ionization mass spectrometry method [8-10], determination of Lafutidine alone and in combination with other drug in Pharmaceutical dosage form [11-13] but as per our literature survey there is no any method have been reported for determination of LAF in human plasma by RP-HPLC using Photo Diode Array detector so here we have developed a rapid sensitive and accurate method for quantitation of Lafutidine in human plasma. As research group is working on analytical development and validation of various bioactive molecule by using sophisticated instrument like HPLC [14,15], so to carry out this work in further direction we have chosen to developed and validate a method to quantify Lafutidine hydrochloride in human plasma.

2. EXPERIMENTAL

2. 1. Materials and Reagents

Pharmacopoeial grade standard of Lafutidine Hydrochloride was received as gift sample by Cadila Pharmaceutical, Ahmedabad. HPLC grade methanol was purchased from Spectrochem Pvt. Ltd., Baroda, water which used in throughout experiment was purified by milli-Q water assembly, analytical grade ammonium acetate was purchased from Spectrochem Pvt. Ltd. and solid phase extraction cartridges (Strata-X) were purchased from Phenomenex, Hyderabad. HPLC column Gemini C18 (4.6 × 250 mm, 5 μm) was purchased from Phenomenex.

2. 2. Instrumentation

The chromatographic system used to perform development and validation of this assay method was comprised of LC-10ATvp binary pump, SPD-M10Avp photo-diode Array detector and Rheodyne manual injector model 7725i with 20 μl loop (Shimadzu, Kyoto, Japan), connected to a multi-instrument data acquisition and data processing system (LC Solution, Shimadzu).

2. 3. Biological samples

Blank human plasma for the development and validation of the method were obtained from government recognise laboratory.

2. 4. Chromatographic Conditions

Chromatographic analysis was performed on a reverse phase HPLC Phenomenex Gemini C18 (4.6 × 250 mm, 5 µm) column, mobile phase consisted of Methanol and 5 mM Di-Potassium Hydrogen Phosphate Buffer (pH 9.5) (80:20, v/v). Flow rate was adjusted to 1.0 ml/min. retention time of analyte was 7.6 min, injection volume was 20 µL and detection was performed at 216 nm.

2. 5. Preparation of stock solution and spiked plasma sample

The stock solution of LAF were prepared in methanol at concentration of 1.0 mg/ml working solution of 50 µg/ml were prepared by appropriately diluting the stock solution of LAF with 50:50 methanol water. Working solution of LAF was used to prepare the spiking stock solution for preparation of nine point calibration curve (50-1000 ng/ml) and quality control (QC) samples at three concentration level LQC (50 ng/mL), MQC (500 ng/mL) and HQC (1000 ng/ml) all stock and working solution were stored in refrigerator (at 2-8 °C) when not in use. The spiked plasma sample was prepared by adding 25 µl of respective spiking stock solution in 475 µl of Blank plasma.

2. 6. Sample preparation and extraction procedure

Extraction was performed by solid phase extraction procedure as described below, 0.5 ml aliquot of the spiked plasma sample was pipetted into 10 ml glass centrifuge tube and mixed with 250 µl of ammonium acetate buffer (pH 4.5), vortexed thoroughly, mixture was loaded in to strata-x cartridge which was pre-conditioned by 1 ml methanol followed by 2 ml water, than cartridge was washed by 1 ml water for two time after that cartridge was transferred into clean test tube and the analyte was eluted with 0.5 ml methanol from which 20 µL volume is injected into HPLC system.

3. METHOD VALIDATION

The validation was executed as per 'Guidance for Industry: Bio analytical Method Validation' from the United States Food and Drug Administration [16].

3. 1. Selectivity

The selectivity of the method was performed to ensure absolute separation of LAF from the biological endogenous components of the human plasma. Selectivity was carried out by analysing seven different lots of blank human plasma by solid phase extraction procedure.

3. 2. Linearity

Five standard curves of nine different concentration standards (50-1000 ng/ml) and two blank samples have been assayed to achieve calibration curve but only nine concentration standards were included in the calibration curve whereas blank sample were used to check interference and contamination.

3. 3. Precision & Accuracy

Intra-day and Inter-day precision and accuracy were evaluated by five spiked samples analyzed at three level of quality control samples (50,500 and 1000 ng/mL) of LAF. Intra -day precession and accuracy was determined by repeated analysis of five spiked samples of LAF at each QC level. Inter-day precession and accuracy were determined by repeated analysis of three consecutive days (5 series per day). The concentration of each sample was determined using standard curves prepared and analysed on same day.

3. 4. Matrix effect

To evaluate the effect of endogenous contribution of different source of blank plasma in the measurement of the analyte the matrix effect was checked. Matrix effect was evaluated by comparing peak area of extracted sample using five different lots of blank human plasma in triplicate at two concentration level of QC (50 and 1000 ng/ml) samples, percentage deviation and relative error were calculated to check interference of matrix effect on the analyte concentration.

3. 5. Recovery

Peak area from unextracted analyte with those of extracted analyte determined recovery. Percentage recovery of LAF was determined by comparing mean area of five replicates each of extracted quality control sample with mean area of freshly prepared unextracted quality control samples which were prepared by spiking blank extracts with respective spiking solution.

4. RESULT AND DISCUSSION

4. 1. Sample preparation

To accelerate drug dissociation of analyte from plasma different buffers were tried, but with 25 mM ammonium acetate (pH 4.5) recovery and response were found better so it was added to plasma sample and methanol was used as eluent.

4. 2. Separation

Different mobile phase were investigated using methanol, acetonitrile and ammonium acetate in various proportions, after several trials Methanol and 5 mM Di- Potassium Hydrogen Phosphate Buffer (pH 9.5) (80:20, v/v) mobile phase was finalised. Flow rate selection was based on peak parameters height, asymmetry, tailing, and baseline drift and run time and set at 1.0 mL/min. The retention time for the investigated drug was found at 7.6 min and runtime was 15 min. Different columns have been tested, with minimal effect on the resolution of the analyte Phenomenex Gemini C18 (4.6 × 250 mm, 5 µm) column have been finalised because of its demonstrated smoothness and reproducibility in this method.

The optimum wavelength for detection of analyte was 216 nm, at which much better detector response was obtained. For the estimation of LAF a sharp and symmetrical peak was obtained with good baseline, which assists the accurate measurement of the peak area.

4. 3. Assay selectivity

Seven blank samples were assayed to examine selectivity of the method therefore it was expected that area of clinical samples would not be prevented by interference peak in this method. Representative chromatogram of blank sample and LLOQ sample are shown in Figure 2. There was no interference of endogenous peak observed at the retention time of LAF results of selectivity are enumerated in Table 1.

Figure 2. HPLC-PDA chromatograms of blank plasma sample and LLOQ (1.0 $\mu\text{g/mL}$) standard (A); chromatogram of extracted blank plasma (B); Chromatogram of extracted LLOQ sample spiked with 1.0 $\mu\text{g/mL}$ of calibration standard.

Table 1. Summary of selectivity experiment results.

Sample name	AREA
LLOQ (50 NG)	285621
LLOQ (50 NG)	276789
LLOQ (50 NG)	279521
LLOQ (50 NG)	258952
LLOQ (50 NG)	290014
MEAN	278179
Blank plasma lot no.	AREA
LOT-1	0
LOT-2	0
LOT-3	0
LOT-4	0
LOT-5	0
LOT-6	0
LOT-7	0
LOT-8	0

4. 4. Linearity

Five calibration curves were assayed as mentioned in section 3.2. The linearity of the method was evaluated by 9 point calibration curve in the concentration range of 50-1000 ng/ml calibration plot is shown in Figure 3.

Figure 3. Calibration curve of Lafutidine Hydrochloride.

All calibration curves were linear over the range of 50-1000 ng/ml. correlation coefficient was ≥ 0.9996 Calibration curve was determined by peak area *versus* concentration.

The mean slope of calibration curves was 5832.75 and mean intercept was -34560.15. Concentration of analyte was calculated by using mean calculation formula $Y = MX + C$ where M = Slope and C = Intercept.

Data of calibration curves is given in Table 2 and Table 3.

Table 2. Back calculated concentration of calibration standards from respective calibration curves.

Linearity	Lafutidine conc. (ng/ml)								
	50	100	200	300	400	500	600	800	1000
1	53.75	102.34	200.30	306.67	381.41	493.15	602.22	793.48	998.77
2	54.98	104.27	201.64	296.22	396.12	495.25	605.83	788.38	1011.97
3	54.35	100.76	198.97	295.40	393.44	490.95	623.15	811.80	979.27
4	52.46	100.74	199.81	297.71	394.99	491.09	615.89	811.18	994.46
5	53.68	99.89	199.81	295.42	395.56	492.80	603.93	801.06	1014.67
Mean	53.84	101.60	200.11	298.29	392.30	492.65	610.21	801.18	999.83
% Mean Accuracy	107.69	101.60	100.05	99.43	98.08	98.53	101.70	100.15	99.98
SD	0.93	1.73	0.98	4.78	6.17	1.76	8.97	10.44	14.32
% CV	1.73	1.71	0.49	1.60	1.57	0.36	1.47	1.30	1.43

Table 3. Parameter corresponding to liner regression obtained from calibration curve of LAF.

Calibration Curve	Slope	Intercept	Correlation Coefficient
1	5787.17	-23116.70	0.9997
2	5827.79	-26340.88	0.9998
3	5819.10	-41906.04	0.9989
4	5843.58	-30890.45	0.9997
5	5886.14	-50546.69	0.9999
Mean	5832.75	-34560.15	0.9996

4. 5. Sensitivity

Sensitivity of the analytical method can be measured by calculating Limit of Detection and Quantification (LOD and LOQ) of analytical method. LOD and LOQ were estimated in accordance with the baseline to noise ratio, it should be 3 and 10 time higher than the blank plasma sample. LOD and LOQ were estimated to be 10 ng/mL and 40 ng/mL respectively, a representative chromatogram of LOD and LOQ is given in Figure 4.

Figure 4. HPLC-PDA chromatogram of LOD and LOQ sample (A); Chromatogram of LOD sample (B); Chromatogram of LOQ sample.

4. 6. Precision and accuracy

As shown in Table 4 Intra-day deviation from nominal concentration was ≤ 4.46 and intra-day precision was ≤ 3.92 . Inter-day deviation from nominal concentration was ≤ 5.12 and inter-day precision was ≤ 3.98 .

Table 4. Intra- and Inter-day precision and accuracy of the method for determination of LAF.

Added conc. (ng/mL)	Intra-Day			Inter-Day ^N		
	Mean calculated conc. (ng/mL)	% RSD	% RE	Mean Calculated conc. (ng/mL)	% RSD	% RE
50	52.23	3.92	4.46	52.29	3.98	4.57
500	499.77	1.32	-0.05	496.89	1.15	-0.62
1000	1003.09	0.86	0.31	1051.19	3.97	5.12

^N For Inter –Day (samples injected were 5 series per day).

4. 7. Matrix effect

Results of matrix effect are given in Table 5. Percentage deviation and relative error of the injected samples at LQC level was $\leq 3.81\%$ and $\leq 7.27\%$ respectively and at MQC level Percentage deviation and relative error was $\leq 3.50\%$ and $\leq 3.26\%$ respectively. Results indicate that there is no considerable endogenous contribution from blank plasma for the measurement of analyte.

Table 5. Result of matrix effect experiment of LAF.

		50 ng/mL			1000 ng/mL		
		Mean	% RSD	% RE	Mean	% RSD	% RE
A-1	3	53.63	1.00	7.27	1021.10	0.95	2.11
A-2	3	51.61	2.22	3.21	990.84	3.50	-0.92
A-3	3	51.60	2.24	3.20	1032.61	0.39	3.26
A-4	3	49.87	3.81	-0.26	1018.45	0.88	1.84
A-5	3	49.77	3.30	-0.46	1012.46	1.79	1.25

4. 8. Extraction recovery

The extraction recovery determined for LAF was consistent, precise and reproducible. As per result given in Table 6. Mean recoveries of the three concentration levels (50, 500 and 1000 ng/mL) were 93.18, 96.08 and 94.45 % respectively, whereas the mean recovery between QC level is 94.57 %.

Table 6. Summary of absolute recovery results of LAF from 0.5 ml of plasma.

	50 ng/mL	500 ng/mL	1000 ng/mL
% Mean Recovery Within QC Level	93.18 %	96.08 %	94.45 %
% RSD	1.56 %	1.51 %	1.54 %
% Mean Recovery Between QC Level	94.57 %		
SD	1.45		
% CV	1.56 %		

4. 9. Dilution integrity

The upper concentration limit of LAF can be extended to 2000 ng/ml with acceptable percentage deviation and relative error of 15 % by 2-fold or 4-fold dilution with blank human plasma. The Results demonstrate percentages deviation and relative error at 2 fold dilution was ≤ 3.59 % and ≤ 2.07 % and for 4 fold dilution deviation and relative error was ≤ 4.29 % and ≤ 2.80 % results are demonstrated in Table 7 which shows that dilution of sample dose not effect on the precession and accuracy of the result obtained by this method so the clinical sample which are above the limit of quantitation can be analysed accurately even if dilution is up to 4 fold.

Table 7. Results obtained by dilution integrity experiment of LAF.

set	Dilution	Dilution factor	Obtain conc.	% RE	% RSD
2000	$\frac{1}{2}$	2	2009.45	0.47	2.07
	$\frac{1}{2}$	2	1928.29	-3.59	
	$\frac{1}{2}$	2	1990.20	-0.49	
	$\frac{1}{2}$	2	2036.94	1.85	
	$\frac{1}{2}$	2	1968.37	-1.58	
2000	$\frac{1}{4}$	4	1914.30	-4.29	2.80
	$\frac{1}{4}$	4	1987.22	-0.64	
	$\frac{1}{4}$	4	2013.10	0.66	
	$\frac{1}{4}$	4	1934.68	-3.27	
	$\frac{1}{4}$	4	2050.79	2.54	

4. 10. Stability

Table 8 summarizes stability experiments which are bench top 24 hour at room temperature, post extraction (at 4-8 °C after extraction), freeze thaw stability(three cycle), long-term stability for 30 days (stored at -20 °C) and post preparative stability data of LAF. All the results showed the stability behaviour during these tests and there were no stability related problems occurred during the validation and stability testing. The stability of working solutions was tested and on the basis of the results these solutions were stable till 6 h.

Table 8. Result of Stability experiment of LAF.

Stability Experiment	QC level	Spiked conc. (ng/mL)	Comparison sample	Stability sample	% Change
Bench top	LQC	50	52.34	52.10	-2.24
	HQC	1000	999.67	1020.88	2.12
Process	LQC	50	51.34	53.23	3.67
	HQC	1000	987.96	995.56	0.77
Freez and Thaw	LQC	50	50.21	52.37	4.31
	HQC	1000	1012.33	995.41	-1.67
Long term	LQC	50	48.54	53.17	9.54
	MQC	500	510.58	502.53	-1.58
	HQC	1000	985.47	1005.52	2.03

Note: Three samples were injected at each level

The bench top stability of LAF was investigated at LQC and HQC levels this revealed that LAF in plasma is stable for at least 6 hrs at room temperature and 24 hrs at auto sampler with mean percentages of $\leq 2.24\%$ and $\leq 2.12\%$ respectively. It was confirmed that repeated freeze and thawing of plasma sample spiked with LAF at LQC and HQC level did not affect the stability of Lafutidine with mean percentage of $\leq 4.31\%$ and $\leq 1.67\%$. Long term stability of the LAF in plasma at -20 °C was also performed for 80 days at LQC, MQC and HQC level with mean percentages change of LAF 9.54 %, 1.58 % and 2.08 %. The results of the stability are enumerated in Table 9. The above results indicate that LAF is stable enough to be analysed using this assay method.

5. CONCLUSION

We have developed a suitable method to quantify Lafutidine hydrochloride in human plasma by solid phase extraction procedure. The method is simple, sensitive, specific and reproducible and may be used in bioavailability and bioequivalence study.

ACKNOWLEDGEMENTS

The authors are thankful of Cadila pharmaceutical Ltd., Ahmedabad For supply us the Lafutidine Hydrochloride, facilities & grants given under UGC- Special Assistance Program Department Research Support (SAP-DRS), Department of Science & Technology (DST) New Delhi, Fund For Improvement of Science & Technology (FIST), National Facility for Drug Discovery and Department of Chemistry, Saurashtra University, Rajkot – 360 005 (INDIA) for providing analytical facilities.

References

- [1] The Merck Index, 2001. An encyclopedia of Chemicals, Drugs and Biologicals, Thirteenth Edition, Merck Research Laboratories division of Merck and Co., Inc. Whitehouse Station, NJ, USA, 1263.2., 5362: 957.
- [2] T. Shimatani, M. Inoue, T. Kuroiwa, J. Xu, M. Nakamura, *Digestive Diseases Science* 51(2006) 114-20.
- [3] H. Sato, K. Kawashima, M. Yuki, H. Kazumori, M. Rumi, C. Ortega, *The Journal of Laboratory and Clinical Medicine* 141 (2003) 102-105.
- [4] H. Toh, T. Aito, M. Akeyama, *Biological Pharmaceutical Bulletin* 25 (2002) 379-382.
- [5] H. Ajioka, N. Matsuura, H. Miyake, *Inflammopharmacology* 10 (2002) 483-493.
- [6] K. Nagahama, M. Yamato, S. Kato, K. Takeuchi, *Journal of Pharmacological Sciences* 93(2003) 55-61.
- [7] H. Iida, M. Inamori, Y. Nozaki, H. Endo, K. Hosono, T. Akiyama, *BMC Gastroenterology* 9 (2009) 52.
- [8] L. Wu, Z. Zhang, Y. Tian, W. Li, F. Xu, Y. Chen, *Journal of Mass Spectrometry* 40 (2005) 1637-1643.
- [9] W. Chen, Y. Liang, H. Li, Y. Xiong, X. Liu, G. Wang, *Journal of Pharmaceutical and Biomedical Analysis* 41 (2006) 256-60.
- [10] K. Jadhav, D. Dhamecha, S. Ghadlinge, G. Asnani, M. Patil, *ISRN Analytical Chemistry* (2012) 1-4.
- [11] M. Sumithra, P. Sundaram, K. Srinivasulu, *International Journal of Chemistry* 3 (2011) 1403-1407.
- [12] M. Jagadeeswaran, N. Gopal, B. Jayakar, T. Sivakumar, *Global Journal of Pharmacology* 6 (2012) 60-64.
- [13] R. Parekh, P. Patel, C. Patel, H. Patel, *International Journal of Drug Development & Research* 4 (2012) 325-329.
- [14] P. Vekariya, H. Joshi, *ISRN Spectroscopy* (2013) 1-6.
- [15] G. Pandya, H. Joshi, *Der Pharmacia Sinica* 4 (2013) 145-152.
- [16] US Department of Health and Human Services, Food and Drug Administration, Center for Drug Evaluation and Research (CDER). Guidance for Industry, Bioanalytical Method Validation, May 2001.