

Study of the Effect of Air to Fuel ratio Parameter on the Organophosphorus – Pesticide Analysis by GC-FPD

A. Y. Kilany, M. A. Elsayed, M. K. Abd Elmegid, M. S. Fayed

Egyptian Armed Forces, Cairo, Egypt
Tel. +201141750544; Fax: +222621918

E-mail address: aboelfotoh@gmail.com

ABSTRACT

In the present contribution, sensitive and precise method for the quantification of Organophosphorus / Pesticides (Malathion and Dimethoate) in nanograms range has been developed. The performance of flame photometric detector (FPD), a selective detector (P&S-mode) that can be used in the analysis of organophosphorus compound, is evaluated in terms of sensitivity, selectivity and reproducibility. The performance of flame photometric detector was strongly depending on the absolute and relative flow rate of air and hydrogen gases. The optimum air-to-fuel ratio for detection of Malathion and Dimethoate was 0.4 and 0.3 (FPD-P mode). At this ratio, low picogram amounts of phosphor can be detected accurately (0.18 pgP) with a wide linear dynamic range of 0.18 pgP to 298 ngP. While, the optimum air-to-fuel ratio, for detection of Malathion and Dimethoate was 0.6 (FPD-S mode). In addition to, the method is precise with 4.5 % relative standard deviation (RSD). In conclusion, it could be proposed that this procedure can be recommended as a suitable method for the quantification of Malathion and Dimethoate in cases of acute poisoning.

Keywords: Sensitivity; Minimum Detectability; Dynamic Range; Linear Range; Organophosphorus pesticides; Malathion; Dimethoate; GC/FPD

1. INTRODUCTION

Organophosphorus (OP) pesticides are widely used in agriculture as insecticides due to their high activity, low bioaccumulation and moderately rapid degradation in the environment. Nevertheless, a certain amount of the pesticides used is transferred in the surface runoff and subsurface drainage from agricultural land. This can cause a spectrum of toxic effects on aquatic organisms and human beings. At the mammalian level, organophosphorus pesticides represent a serious risk because of their irreversible binding to the active site of acetyl cholinesterase (AChE). This leads to hindered hydrolysis of the acetylcholine (AChE) resulting in the dysfunction of the transmission of nerve impulses with poisoning effects and nervous diseases. The measurements of OP compounds are classically done using chromatographic techniques [1-6].

Malathion S-[1,2-di(ethoxycarbonyl)ethyl] dimethyl phosphorothiolothionate is a member of a general class of organophosphorus or organophosphate (OP) pesticides. Malathion is not only used to control insects on fruits and vegetables, but is also employed in

2. EXPERIMENTAL

2. 1. Reagents and solutions

Malathion and Dimethoate were purchased from Nasr Company for intermediate chemicals, Egypt with purity > 95 %. Stock standard solutions were prepared by accurately weighing 0.3 g of pure material. Malathion and Dimethoate were dissolved in pesticide-quality benzene and diluted to a volume of 25 mL in volumetric flask. Stock standard solutions were transferred into TFE-fluorocarbon-sealed screw-cap vials, stored at 4 °C and protected from light. A set of test samples were prepared by dilution from stock standard solutions, covering a wide range of concentrations. A fixed volume (1 µL) of each concentration of the test substance was injected, the height (H) and integrated peak area (A) of the resultant Gaussian or chromatographic peaks were measured. Frequently stock standard solutions were checked for signs of degradation or evaporation.

2. 2. Instrumentation and analysis conditions

Agilent, 7890A gas chromatograph, Autosampler (7693-series) with a split/ Splitless injector system and a flame photometric detection with phosphorus filter were used. Ultra pure nitrogen (99.9999 %) at 25 psi constant pressure was passed through a molecular sieve trap and an oxygen trap was used as the carrier gas. The injection port was held at 200 °C and used in the splitless mode. Separation was carried out on a HB-5, 30 m x 0.32 mm capillary column with a 0.25 µm stationary film thickness (Agilent Technologies). Balance: Analytical, capable of accurately weighing to the nearest 0.0001 g. The column temperature was maintained at 60 °C for 1 min and then programmed at 20 °C /min to 250 °C, and held for 5 min. The total analysis time was 15.5 min. The volume of sample injected in splitless mode was 1.00 µL. Absolute and relative flow rate of air and hydrogen used for flame photometric detector were studied from 0.2-1.2 air-to-hydrogen ratio.

3. RESULTS AND DISCUSSIONS

3. 1. Influence of air-to-hydrogen flow rates ratio on FPD performance:

The intensity and wavelength of light emitted from the FPD flame depends on the geometric configuration of the flame burner and on the absolute and relative flow rates of gases supplied to the detector. By judicious selection of burner geometry and gas flow rates, the FPD flame is usually designed to selectively enhance optical emissions from certain types of molecules while suppressing emissions from other molecules [11]. Figures 3 and 4 show the effect of relative air-to-hydrogen ratio flow rates on FPD performance at constant organophosphorus concentration (1µL injection). The result show that the optimum condition for detector response with malathion is 0.4 while for dimethoate is 0.3. The make-up gas flow rate was found to have only a marginal effect on the FPD performance. In all further experiments a make-up flow of 60 ml/min was used.

The detector temperature was set at 250 °C. The sensitivity of the detector was found to be maximum at 0.4 air-to-fuel ratio for malathion and 0.3 with dimethoate. The chromatograms of detector responses against air-to-H₂ ratio for malathion and dimethoate are shown in Figure 5 and 6 as 3D figure, respectively.




Fig. 3. Detector responses at different air-to-hydrogen ratio GC-FPD-P mode for Malathion, measured at GC operating condition.




Fig. 4. Detector responses at different air-to-hydrogen ratio GC-FPD-P mode for Dimethoate, measured at GC operating condition.




Fig. 5. 3D chromatograms of different air to H₂ ratio GC-FPD-P mode for Malathion, measured at GC operating condition.




Fig. 6. 3D chromatograms of different air to H₂ ratio GC-FPD-P mode for Dimethoate, measured at GC operating condition.

3. 2. Effect of total gases flow rates in the performance of FPD-P mode

In this type of measurements, the total gas flow rate was varied from 105 to 210 ml/min at air to hydrogen ratio 0.4 for Malathion and from 65 to 195 ml/min at air to hydrogen ratio 0.3 for Dimethoate. Detector responses were determined at different total gas flow rate and the results are shown in Figures 7 and 8 for both Malathion and Dimethoate, respectively. Figures 9 and 10 shows the 3D chromatograms of detector responses verses total gas flow rates for Malathion and Dimethoate, respectively. As indicated from the figures, the performance increases with increasing total flow rates.

This can be attributed to Hydrogen – Air or Hydrogen – Oxygen diffusion flames are normally employed for the FPD. In such diffusion flames, the hydrogen and oxygen do not mix instantaneously, so that these flames are characterized by significant spatial variations in both temperature and chemical species. The important chemical species in hydrogen – air flame are the H, O, and OH flame radicals.

These highly reactive species play a major role in decomposing incoming samples and in the subsequent production of the desired optical emissions. Optical emissions from the HPO molecular systems are highly favored in those regions of an FPD flame which are locally rich in H-atoms, while CH and C₂ light emissions from hydrocarbons originate mainly from those flame regions which are locally rich in O-atoms.

The highest sensitivity and specificity for phosphorus detection are achieved only when the FPD flame is operated with hydrogen in excess of that stoichiometric amount required for complete combustion of the oxygen supplied to the flame. This assures a large flame volume that is locally abundant in H-atoms, and a minimal flame volume that is locally abundant in O-atoms, [11].




Fig. 7. Detector responses versus total gases flow rate for Malathion GC-FPD-P-mode, at the same operating condition.




Fig. 8. Detector responses versus total gases flow rate for Dimethoate GC-PFD-P mode, at the same operating condition.




Fig. 9. (3D) Chromatograms of detector responses versus total gases flow rates for Malathion, at the same operating condition.




Fig. 10. (3D) Chromatograms of different total gases flow rate for Dimethoate, at the same operating condition.

3. 3. Linearity of FPD-P mode with Malathion and Dimethoate




Fig. 11. Linear range of the detector (FPD-P-mode) for Malathion measured at optimum operating conditions.

The linearity of the detector FPD-P mode was determined by the injection of 1 μ L standard solution at different concentrations. Calibration curves plotted from different concentrations were obtained for analytes by regression analysis of peak area versus injection concentration.

Figures 11 and 12 shows the relation between detector response (area under the curve) and concentrations of pesticide. The plot shows the linearity of the detector with Malathion and Dimethoate with R^2 value higher than 0.997 over the studied concentrations range, which indicate good linearity. Similar results for R^2 (0.999-0.9823) were obtained in a study using GC/PFPD and GC/MS for the determination of trace OPPs, [12-14].

Figures 11 and 12 presented the determined linearity of the detector with Malathion and Dimethoate, respectively. The linear range of the detector for Malathion covers the concentration range from 40.2×10^{-3} to 3175.819 mg/l with correlation coefficient value (R^2) = 97.6% while it was from 3.8 to 638.005 mg/l with correlation coefficient value (R^2) = 97.2% for Dimethoate




Fig. 12. Linear range of the detector (FPD-P-mode) for Dimethoate measured at optimum operating conditions.

3. 4. Accuracy and precision measurement of the method for Malathion and Dimethoate

The precision and accuracy of the methods were evaluated using relative standard deviation for replicates ($n = 5$). Standard concentrations of 1385.273 and 315.439 mg/l for Malathion and Dimethoate, respectively were injected 5 times, the results are shown in Table 1 and 2. The relative standard deviation of the peak areas of the pesticides were (RSD) = 1.979 % and 0.363 % for Malathion and Dimethoate, respectively. The result revealed that, the deviation was not exceeding 20 %, in accordance with the methods of validation of pesticide residues reported in other research [15]. In addition to, the deviation was in a range similar or better than those obtained in other studies using GC-FPD-P mode (1.7, 1.2, and 5.6 %) [16].

Table 1. Five injection of Malathion in GC-FPD-P-mode.

Sample no.	Area (150pA/s)	Conc. according to calibration curve (mg/l)
1	1056032.4	1625.602
2	1051107.6	1617.313
3	1060550.9	1633.208
4	1062909.7	1637.178
5	966233.9	1474.452
Mean	1039366.9	1597.551
Accuracy		15.3 %
Precision		1.979%

Table 2. Five injection of Dimethoate in GC-FPD-P-mode.

Sample no.	Area (150pA/s)	Conc. according to calibration curve
1	823896.1	341.6105
2	825796.1	342.2671
3	830050.6	344.3853
4	825316.7	342.2509
5	827216.7	343.1076
Mean		342.7243
Accuracy		8.6499%
Precision		0.363%

3. 5. Determination of Limit Of Detection (LOD) and Limit Of Quantification (LOQ) for the detector GC-FPD-P mode for Malathion and Dimethoate

LOD and LOQ are measured for the methods by the analysis of ultra-pure samples (Malathion and Dimethoate) with decreasing concentrations until reach the lowest detectable level of the organophosphorus pesticides that could be detected by the GC-FPD-P mode.

The values of LOD were found to be 1.88×10^{-3} mg/l and 1.52 mg/l for Malathion and Dimethoate, respectively these concentrations equal to three times the signal/noise ratio. On the other hand, the lowest determinable levels (LOQ) were 0.017 mg/l and 5.7 mg/l for Malathion and Dimethoate respectively, these concentrations equal to ten times the signal/noise ratio. Figure 13 and 14 shows LOD and LOQ for Malathion. The method show good sensitivity for the detection of Malathion compared to Dimethoate, this may be due to the differences in the chemical and physical properties of both compounds and their interaction with the stationary phase of the column.




Fig. 13. Chromatogram of 1.88×10^{-3} mg/l Malathion (LOD), at optimized conditions.




Fig. 14. Chromatogram of 0.017 mg/l Malathion (LOQ), at optimized condition

The samples components are usually held on the stationary phase so strongly that they are eluted very slowly or even not at all [6]. In this method, the used column is HP-5 with non polar stationary phase which is favor the non polar compounds such as Malathion, while it is not favor the compounds have polarity such as Dimethoate.

4. CONCLUSIONS

In conclusion, the study of air-to-fuel ratio with flame photometric detector for the analysis of Malathion and Dimethoate samples showed the validity of Gas Chromatography-Flame Photometric Detector (P and S -modes) providing excellent separation and peak shape. From the results carried out for the analysis in this study, it was revealed that, the method applied using GC-FPD, the optimum air to hydrogen ratio is at regions of the FPD flame which are locally rich in H-atoms such as 0.4 for the analysis of Malathion and 0.3 for the analysis of Dimethoate with P mode and 0.6 with S-mode. The highest sensitivity and specificity for phosphorus detection are achieved only when the FPD flame is operated with hydrogen in excess of that stoichiometric amount required for complete combustion of the oxygen supplied to the flame. The method is precise, reproducible, simple, rapid, has a wide range of linearity and dynamic range over a wide range of analyte concentrations with correlation coefficient ($R^2 = 0.968$) and (RSD) < 4.5 %, and uses small volumes of solvents and samples, which reducing the risk for human health and the environment. Besides, the results of this study show that the proposed method could be further applicable to determine the residues of different pesticides in environment.

References

- [1] Bowman M., M. Beroza, *Journal of Chromatographic Science* 7(8) (1969) 484-492.
- [2] Crider W. L., *Analytical Chemistry* 41(3) (1969) 534-537.
- [3] Bennett G. F., *Introduction to Environmental Analysis*: Roger Reeve, John Wiley & Sons, Inc., New York, NY, 2002, Elsevier.
- [4] Brody S. S., J. E. Chaney, *Journal of Chromatographic Science* 1966. 4(2) (1966) 42-46.
- [5] Versino B., G. Rossi, *Chromatographia* 4(8) (1971) 331-336.
- [6] Dagnall R., K. Thompson, T. West, *Analyst* 93(1103) (1968) 72-78.
- [7] Marek B., Todd S., Aamand S., *Journal of Enzyme Inhibition and Medicinal Chemistry* 18(6) (2003) 551-555.
- [8] WHO, "Malathion in Drinking water", 2004, WHO/SDE/WSH/103.
- [9] A. Ayman, Gouda Alaa S., Amin Ragaa El Sheikh, *Chemical Industry & Chemical Engineering Quarterly* 16(1) (2010) 11-18.
- [10] WHO. Environmental Health Criteria monograph "Dimethoate", WHO, Geneva, 1989.
- [11] Dömötöróvá M., E. Matisová, *Journal of Chromatography A* 1207(1) (2008) 1-16.
- [12] Wang D., D. P. Weston, M. J. Lydy, *Talanta* 78(4) (2009) 1345-1351.
- [13] Saeed T., et al., *Food Control* 12(2) (2001) 91-98.

- [14] Frishman G., A. Amirav, *Field Analytical Chemistry & Technology* 4(4) (2000) 170-194.
- [15] Bidari A., et al., *Food Chemistry* 126(4) (2011) 1840-1844.
- [16] Samadi, S., H. Sereshti, and Y. Assadi, *Journal of Chromatography A* 1219 (2012) 61-65.

(Received 21 June 2014; accepted 01 July 2014)